



महाराष्ट्र शासन

शासकीय वैद्यकीय महाविद्यालय व सर्वोपचार रुग्णालय बारामती
Government Medical College & General Hospital Baramati

प्लॉट पी-१०७, एमआयडीसी एरिया, महिला रुग्णालयासमोर, बारामती जि.पुणे ४१३१३३

Email- deangmcbaramati@gmail.com

जा.क्र.शावैमज/खरेदी विभाग/यंसाखरेदी/दरपत्रक/ १९४१

/२०१९

दि.२५/११/२०१९

दरपत्रक मागणी पत्र :-

प्रति

विषय :-शासकीय वैद्यकीय महाविद्यालय व सर्वोपचार रुग्णालय बारामती या संस्थेतील विविध विभागासाठी यंत्रसामुग्री खरेदी करण्यासाठी दरपत्रके सादर करणेबाबत.

संदर्भ:- शासन निर्णय क्र.जीसीबी-२०१९/प्र.क्र.१६५/प्रशा-१, दिनांक २४ जुन, २०१९
अन्वये प्रशासकीय मान्यता प्राप्त.

शासन निर्णय दिनांक ०५ जुलै, २०१९ अन्वये बारामती जि.पुणे येथे शैक्षणिक वर्ष २०१९-२० पासून १०० विद्यार्थी प्रवेशास परवानगी मिळाली असून प्रत्यक्षात महाविद्यालय सुरु झालेले आहे. महाविद्यालयाच्या अधिपत्याखाली विभागाकरिता वित्तीय वर्ष २०१९-२० करिता यंत्रसामुग्री व साधनसामुग्री खरेदी करण्यासाठी इच्छुक पुरवठाधारकांकडून मोहोरबंद पाकीटात दरपत्रके मागविण्यात येत आहे. तरी इच्छुक पुरवठाधारकांनी त्यांचे दरपत्रके मोहोरबंद पाकीटात खाली नमुद केलेल्या यंत्रसामुग्री व साधनसामुग्रीकरिता सोबत जोडलेल्या विनिर्दिष्टानुसार दरपत्रके दि.०५/१२/२०१९ पर्यंत सादर करावी. सदरची दरपत्रके शासकीय वैद्यकीय महाविद्यालय व सर्वोपचार रुग्णालय बारामती प्लॉट पी-१०७, एमआयडीसी एरिया, महिला रुग्णालयासमोर बारामती जि.पुणे ४१३१३३ येथे कार्यालयीन वेळेत सकाळी ११:०० ते ५:०० पर्यंत या कार्यालयाच्या आवक शाखेत खालील अटी व शर्तीनुसार सादर करावी.


अ.क्र.	विभागाचे नाव	यंत्राचे नाव	यंत्राचे विनिर्दिष्टये (Specifications)
सोबत यादी जोडलेली आहे.			

महत्वाच्या अटी व शर्ती

- १) दरपत्रक मोहोरबंद लिफाफा दि.०५/१२/२०१९ रोजी सायं ५.०० वाजेपर्यंत कार्यालयास पोहचतील या वेताने पाठवावे. जाहीर केलेल्या तारखेनंतर अथवा कार्यालयीन वेळेनंतर आलेले दरपत्रक विचारात घेतले जाणार नाही / ग्राह्य धरले जाणार नाही.
- २) सूचना फलक तसेच संस्थेच्या www.gmcbaramati.org या संकेतस्थळावर जाहीर नोटीसप्रमाणे इच्छुक कंपनीस देखील दरपत्रक मोहोरबंद लिफाफा सादर करण्याची मुभा देण्यात येत आहे.
- ३) यंत्राची किंमत लागू असलेल्या (GST) सर्व करांसहित, संबंधित विभागामध्ये (Door Delivery), वस्तुच्या स्थापत्यासहित एकूण किंमत नमुद करावी.
- ४) प्रत्येक विभागानुसार यंत्रसामुग्रीसाठी स्वतंत्र दरपत्रक मोहोरबंद पाकीटात सादर करावे. म्हणजेच एका पाकीटात एका विभागाच्या सर्व यंत्रसामुग्रीचे दर नमुद करून दरपत्रक सादर करण्यात यावे. अन्यथा दरपत्रक अग्राह्य धरण्यात येईल.

- ५) दरपत्रकाच्या पाकिटावर या कार्यालयाचा पत्र क्रमांक, दिनांक, विभागाचे नाव, यंत्रसामुग्रीचे नाव, अॅटम क्रमांक नमुद करण्यात यावे.
- ६) अधिकृत लेटरहेड वर दरपत्रक सादर करावे. सदरच्या लेटरहेडवर GST नोंदणी क्रमांक, PAN क्रमांक असणे आवश्यक आहे.
- ७) यंत्रसामुग्रीसाठी दरपत्रकात आकारण्यात आलेले दर हे बाजारभावापेक्षा अधिक नसावेत. सर्वात कमीत कमी दराचे पत्रक व वस्तुचा दर्जा/गुणवत्ता उत्कृष्ट असेल, अशाच दरपत्रकाचा विचार करण्यात येईल.
- ८) दरपत्रक ग्राह्य कालावधी हा पुरवठा आदेश नोंदविलेल्या दिनांकांपासून सहा (६) महिन्यांचा राहिल. दरपत्रकात खाडाखोड करण्यात येऊ नये.
- ९) मागणी करण्यात आलेली वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री ही विनिर्दिष्टानुसार व उत्कृष्ट दर्जाची व गुणवत्तेची असणे अत्यंत आवश्यक आहे. विनिर्दिष्टानुसार असलेल्या यंत्रसामुग्रीचेच दरपत्रक सादर करण्यात यावे. सदर दरपत्रकासोबत यंत्रसामुग्रीनुसार विनिर्दिष्टे विहित नमुन्यात नमुद करण्यात यावी किंवा जोडण्यात यावे.
- १०) १) उत्पादक कंपनीचे ऑथोरायझेन प्रमाणपत्र २) वस्तु/यंत्रसामुग्री/उपकरण/ साधनसामुग्रीचे माहितीपत्रक, कॅटलॉग/ब्राऊचर्स यांची मुळ प्रत, ३) आय.एस.ओ./आय.एस.आय. प्रमाणपत्र, ४) पुरवठादार हा अधिकृत विक्रेता असल्याचे संबंधित कंपनीचे अधिष्ठाता यांचे नावे असलेले प्रमाणपत्र ५) पुरवठा करण्यात येणाऱ्या वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री यांचा वॉरंटी कालावधीत संपल्यानंतर पुढील आठ वर्ष प्रतिवर्षी संस्थेच्या आवश्यकतेनुसार वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री खरेदीच्या १ टक्के प्रमाणे वार्षिक देखभाल दुरुस्ती व सेवा करार अथवा ५ टक्के प्रमाणे सर्व समावेशक देखभाल दुरुस्ती सेवा करार करण्यास दरपत्रकधारकास बंधनकारक राहिल असे हमीपत्र जोडावे (सोबत हमीपत्राची प्रत संलग्न),
- ११) यंत्राचे प्रात्यक्षिक (Demo) :- खरेदी करावयाच्या यंत्रसामुग्रीचे संबंधित विभागाचे अध्यापक तथा विभागप्रमुख तथा तांत्रिक सदस्य यांना प्रात्यक्षिक देणे बंधनकारक राहिल.
- १२) बरील अट क्रमांक १० मध्ये नमुद अ.क्र.१, २, ३, ४, ५ तसेच अट क्रमांक ११ ही दस्तावेज ही लहान वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री यांना आवश्यकतेनुसार शिथिल करण्याचे अधिकार अधिष्ठाता यांना राहतील.
- १३) यंत्राचा हमी कालावधी (Warranty Period) यंत्राचा हमी कालावधीत यंत्र समाधानकारकरित्या कार्यान्वित करण्यात आलेल्या दिनांकापासून पुढे दोन (२) वर्षांचा राहिल. हमी कालावधीत यंत्रात तांत्रिक समस्या उद्भवल्यास तात्काळ निराकरण करणे पुरवठादावर बंधनकारक राहिल. अन्यथा हमी कालावधीत वाढ करण्यात येईल किंवा Replacement द्यावे लागेल.
- १४) वार्षिक देखभाल दुरुस्ती करारनामा (AMC & CMC Agreement) :- यशस्वी दरपत्रकधारकास पुरवठा आदेशातील अटी व शर्तीनुसार वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री देखभाल व दुरुस्ती वार्षिक करारपत्र रु.१००/- च्या स्टॅम्पेपरवर (नोटरीसह) स्वतंत्रपणे कार्यालयास सादर करावा लागेल. प्रतिवर्षी विभागप्रमुखांच्या यंत्राच्या देखभाल दुरुस्तीविषयीचा समाधानकारक अहवाल प्राप्त झाल्यानंतरच देयकाची अदायगी कार्यवाही करण्यात येईल.
- १५) यंत्राचा पुरवठा (Delivery):- पुरवठा आदेश निर्गमित झाल्याच्या दिनांकापासून (३०) दिवसांत वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्रीचा पुरवठा करण्यात यावा. विहित मुदतीत पुरवठा करून वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री प्रस्थापित व कार्यान्वित न केल्यास होणाऱ्या विलंबास नियमानुसार दंडात्मक कार्यवाही करण्यात येईल.

- १६) देयकाची अदायगी (Payment):- पुरवठा करण्यात आलेले वस्तु/यंत्रसामुग्री/उपकरण/ साधनसामुग्री समाधानकारकरित्या प्रस्थापित व कार्यान्वित करून दिल्यानंतर, विनिर्दिष्टानुसार व उत्कृष्ट दर्जाचे/गुणवत्तेची असल्याबाबत संबंधित विभाग तथा संस्थेतील तांत्रिक समिती यांचेकडून स्वतंत्र अहवाल कार्यालयास प्राप्त झाल्यानंतरच ९० टक्के रक्कम अदायगी करण्याची कार्यवाही करण्यात येईल तसेच उर्वरित १० टक्के रक्कम वस्तु/यंत्रसामुग्री/उपकरण/साधनसामुग्री कार्यान्वित झाल्यापासून तीस (३०) दिवसांच्या कालावधीपर्यंत संबंधित विभागाने यंत्राचे उपयोगिता प्रमाणपत्र प्राप्त झाल्यानंतर तसेच पुरवठा आदेशातील सर्व अटी व शर्तीनुसार संपुर्ण दस्तऐवजांची पूर्तता केल्यानंतरच उर्वरित रक्कम अदा करणेत येईल.
- १७) प्रशिक्षण (Training) आणि यंत्राचे श्रेणीवर्धन (Upgrades) :- यशस्वी दरपत्रकधारकास विभागातील किमान दोन अधिकारी व व कर्मचाऱ्यांना सदर वस्तु/यंत्रसामुग्री/उपकरण/ साधनसामुग्री निःशुल्क प्रशिक्षण (Training) देणे बंधनकारक राहिल. तसेच साफ्टवेअरमध्ये वेळोवेळी बदल झाल्यास ते निःशुल्क पुरविणे बंधनकारक राहिल.
- १८) यंत्राचा पुरवठा आदेश दिल्यानंतर दरपत्रक तसेच पुरवठा आदेशातील अटी व शर्तीनुसार यंत्राचा पुरवठा विहित मुदतीत करणे बंधनकारक राहिल. विहित मुदतीत पुरवठा न केल्यास द्वितीय न्यूनतम दरपत्रकधारकास पुरवठा आदेश देण्यात येतील.
- १९) अपवादात्मक परिस्थितीमुळे उक्त कोणत्याही अटीत आवश्यकतेनुसार अंशतः बदल करणे, अथवा प्राप्त झालेली दरपत्रके कोणतेही कारण न देता स्विकारणे/नाकारणे तसेच कोणत्याही स्तरावर खरेदीप्रक्रिया रद्द करणे इ. अधिकार संस्थाप्रमुख (अधिष्ठाता) यांनी राखून ठेवलेले आहेत.



अधिष्ठाता

शासकीय वैद्यकीय महाविद्यालय,
बारामती

प्रत:-संस्थेचे सूचना फलक / संकेतस्थळ जाहीर नोटीस शा.वै.म.बारामती.
प्रत :-सर्व संबंधित प्राध्यापक व विभागप्रमुख यांना माहितीस्तव.

दरपत्रक धारक अधिकृत विक्रेता/पुरवठादार/कंपनी यांचे लेटरहेड

Letter No.

Date :-

प्रति

अधिष्ठाता,

शासकीय वैद्यकीय महाविद्यालय,

बारामती.

विषय :- यंत्राचा वार्षिक देखभाल दुरुस्ती सेवा करार (AMC) सर्वसमावेशक देखभाल दुरुस्ती सेवा करार (CMC) विषयी हमीपत्र...(अट क्र. १४ नुसार)

संदर्भ :- आपले दरपत्रक मागणी पत्र

क्रमांक.

दिनांक

महोदय

संदर्भाकित दरपत्रक मागणीपत्रानुसार मागणी करण्यात आलेले यंत्राचा वार्षिक देखभाल करार (AMC) सर्वसमावेशक देखभाल दुरुस्ती सेवा करार (CMC) करणे विषयी सादर करण्यासंबंधीचे हमीपत्र मी खालीलप्रमाणे देत आहे.

हमीपत्र (Undertaking)

याद्वारे हमी देतो की, सन २०१९-२० या वित्तीय वर्षात आमच्याद्वारे पुरवठा करण्यात आलेल्या यंत्रसामुग्रीच्या देखभाल दुरुस्तीस (यंत्राच्या हमी कालावधीसह पुढिल ८ वर्ष) आम्ही बांधील आहोत.

यंत्रांच्या हमी (Warranty) कालावधीत कोणतीही तांत्रिक समस्या निर्माण झाल्यास अथवा यंत्रे संपुर्णपणे बंद पडल्यास २४ तासात यंत्राच्या समस्येचे निराकरण करून देण्यात येईल अथवा नविन यंत्रांचा पुरवठा करण्यात येईल. यासाठी कोणतेही अतिरिक्त शुल्क आकारण्यात येणार नाही. तशी मागणी करण्यात येणार नाही. यंत्राचा २ वर्षांचा हमी कालावधीनंतर (Warranty Period)पुढिल ८ वर्ष प्रतिवर्षी संस्थेच्या आवश्यकतेनुसार यंत्राच्या खरेदी किंमतीच्या १ %प्रमाणे वार्षिक देखभाल दुरुस्ती सेवा करार (AMC) अथवा ५ %प्रमाणे सर्वसमावेशक देखभाल दुरुस्ती सेवा करार (CMC) करणे आम्हास बंधनकारक राहिल. त्यासाठी अटी व शर्तीप्रमाणे करारपत्रे वळोवेळी कार्यालयास सादर करण्यात तयार आहे.

खरेदी करार (Purchase Agreement), वार्षिक सर्वसमावेशक देखभाल दुरुस्ती करार (AMC & CMC Agreement) यांचा भंग झाल्यास नियमानुसार होणाऱ्या कार्यवाहीस मी/आम्ही पात्र असू. तद्संबंधी माझी/आमची कुठलीही हरकत राहणार नाही. अशी याद्वारे हमी देत आहे.

अधिकृत व्यक्तीचे नाव व स्वाक्षरी,
शिक्का अणि मोहर
Authorized Person Name,
Signature
with Stamp and Seal

अस्थिव्यंगोपचारशास्त्र विभाग

अॅटम क्रमांक	यंत्राचे नाव	आवश्यक नग संख्या
२	हायड्रोलिक ऑपरेशन टेबल विथ ऑर्थोपेडीक अॅटॅचमेटस्	१

स्त्री-रोग व प्रसुतीशास्त्र

अॅटम क्रमांक	यंत्राचे नाव	आवश्यक नग संख्या
५	एल. एस. सी. एस. सेट	३
६	हायस्टेरेक्टोमी सेट	३
७	एम.टी.पी. सेट	३
८	लॅप्रोस्कोपीक टयुबेक्टोमी	५

औषधवैद्यकशास्त्र

अॅटम क्रमांक	यंत्राचे नाव	आवश्यक नग संख्या
९	बेड साईड स्टुल	१००
१०	बेड साईड ट्रॉली	१००
११	एक्झामिनेशन टेबल	५

शल्यचिकित्साशास्त्र

अॅटम क्रमांक	यंत्राचे नाव	आवश्यक नग संख्या
१२	सर्जिकल ड्रम ९ बाय ११	१००

बधिरिकरणशास्त्र

अॅटम क्रमांक	यंत्राचे नाव	आवश्यक नग संख्या
१४	एअर कंडीशनर २ टन	३



DEAN

GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE

item NO-2

Hydraulic Operation Table with Orthopedic Attachment

Highly maneuverable Table on Heavy duty Hydraulic Pump ✓

Table base and all accessories made from 304 grade Stainless steel ✓

Accessories made from SS 304 grade ✓

Castor wheels with ball bearings for easy mobility ✓

Lever controlled lock for castors ✓

Table top size:- 485mm x1800 mm

Elevation 800mm – 1000mm / (1250mm)

Trendlenberg 30 Degrees on each side.

Lateral tilt:- 20 degrees on each side.

Back brake:- 90 degrees

Weight 250 kgs aprox. ✓

Accessories

Mattress:- 1 set

Wristlets:- 1 set

Padded Thy rest:- 1 set

Shoulder rest :- 1 set

Intravenous arm board :- 1 set

Anesthetic screen:- 1 no

Ortho Attachments for Table

OR 1 TOTAL HIP REPLACEMENT (THR)

OR 2 HAND SURGERY

OR 3 CENTRAL POST

OR 4 TIBIA NAILING

OR 5 FEMUR NAILING

Aswade
PROFESSOR AND HEAD
Dept. of Orthopedics
Govt. Medical College and General Hospital
Baramati, Dist- Pune

Aswade
DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE

Item No-5
L.S.C.S INSTRUMENTS SET

ITEM DETAILS:

- 1 Sponge Holding Forceps 25cm
- 4 Green Armytage Forceps
- 6 Artery Forceps cvd 15cm
- 6 Artery Forceps St 15cm
- 4 Allice Tissue Forceps Medium 15cm
- 4 Allice Tissue Forceps Large 15cm
- 2 Beacock Tissue Forceps 15cm
- 2 Toothed Forceps 15 cm
- 2 Dissecting Forceps Toothed 15cm
- 2 Dissecting Forceps Non Toothed 15cm
- 1 Needle Holder Mayo Heger 15 cm
- 4 Kell's Clamp 15cm
- 1 Suction Tip
- 1 Tissue Cutting Scissors 17.5cm
- 2 B.P.Handle no4
- 4 Towel Clip Cross Action
- 1 Bowl 20"
- 1 Doyen Retractor 5 cm
- 1 Morr's Retractor 5cm
- 1 Self Retaining Retractor
- 1 Dever's Retractor 5 cm
- 1 Instruments Bag

1 Suture cutting scissors


DEAN

GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE



PROFESSOR AND HEAD
Dept. of Obstetrics and Gynaecology
Govt. Medical College and General Hospital
Baramati, Dist- Pune

Item No- 6 Hysterectomy set

Instrument Set Contents

- 2 Foerster Sponge Forceps Straight 9-1/2"
- 2 Allis Tissue Forceps 5x6 Teeth 9-1/2"
- 2 Mixer Rt Angle Forceps.
- 2 Curved Hemostat.
- 2 Heaney Needle Holders.
- 6 Towal Clip.

1 B.P. handle.
6 Artery forceps curved
2 Artery forceps straight

Shankar
DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE

4 Allis tissue forceps medium

4 Allis forceps large

2 Babcock tissue forceps

4 Kocher's clamp curved

2 Kocher's clamp straight

2 Tissue cutting scissors

1 Mayers extractor

1 self retaining extractor

1 Right angle extractor.

Shankar

PROFESSOR AND HEAD
Dept. of Obstetrics and Gynaecology
Govt. Medical College and General Hospital
Baramati, Dist- Pune

Item No-7

Gynecology Instruments. M.T.P Set.

1 Ovum Forceps. ...

1 Rochster Pean Forceps. ...

1 A V Retractors.

1 Sims Speculum

1 Dialeter set

1 Cusco Speculum

1 vulsellum .

1 Suction cannula set.


2 MVA syringe.

1urette (both sharp & blunt ends)



DEAN

GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE



PROFESSOR AND HEAD

Dept. of Obstetrics and Gynaecology
Govt. Medical College and General Hospital
Baramati, Dist- Pune


Item No-8

TUBECTOMY INSTRUMENT SET

Each Set Contents	Quantity
Sponge Holding Forceps 10"	1
Needle Holder 8"	1
Tissue Forceps Allice	2
Artery Forceps 6" Str	1
Artery Forceps 6" Cvd	1
Bebcock SS 15cm	1
Towel Clip 10cm	1
Scissors Sharp/ Sharp 15cm	1
Scissors Sharp/ Blunt 15cm	1
Scissors Curved 15cm	1
Abdominal Retractor Langenback	1
Vaginal Retractor	1
Cuscos Speculam Large	1
B.P Handle No.4	1
Blade of 6 pkt	1
Suture Needle Str	1
Dissecting Forceps Toothed 6"	1



DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE



PROFESSOR AND HEAD
Dept. of Obstetrics and Gynecology
Govt. Medical College and General Hospital
Baramati, Dist. Pune

Item No-9

BED SIDE STOOLS

- A. Overall Size: Size: 300mm X 300 X 490mm ht./12" X 12" X 19" HT.
- B. Tubular structure 25.4 x 25.4 x 1.25 mm thick ERW tube.
- C. Legs fitted with nylon feet with Metal inserts
- D. Top made form 1 mm Thick CRCA Sheet with a separate frame work to fix as well as to support the top.
- E. All welded joints to finished and polished
- F. Finish: All components should be pretreated in separate eight tank process for better finish, good adhesion and corrosion protection. Process includes Hot Degreasing, De rusting, Activation, Phosphating & Number of Water rinses as per IS 3618 - 1966 class 'C' type and then pretreated materials is coated with epoxy powder with film thickness of 60 microns (approx.) and then oven baked at 180 degree centigrade

Dr. J. S. Patil

PROFESSOR AND HEAD
Dept. of General Medicine
Govt. Medical College and General Hospital
Baramati, Dist- Pune

Dr. S. S. Patil
DEAN

**GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE**

Item No-10

BED SIDE TROLLEY

- a) Overall Size: 430 mm L x 300 mm W x Ht 810 mm(Approx.) /17" X12" X 32" ht
- b) Tubular structure. All vertical sections made form 19 x 19 mm x 18 swg tube
- c) Top and Shelves should to be made from 1mm thick CRCA sheet press bent on all 4 sides.
- d) Top, middle and bottom shelf to be provided with railings on three sides made from 12mm dia S.S 304 grade tube
- e) Provided with 50 mm non rusting & non marking revolving castors made form high strength engineering plastic
- f) Top and shelves is matt finish to prevent slippage hands and items kept on shelves and resist form scratches.
- g) Completely seamless weld at joints
- h) Load bearing capacity should be at least 25 kg.
- i) Finish: Finish: All components should be pretreated in separate eight tank process for better finish, good adhesion and corrosion protection. Process includes Hot Degreasing, De rusting, Activation, Phosphating & Number of Water rinses as per IS 3618 - 1966 class 'C' type and then pretreated materials is coated with epoxy powder with film thickness of 60 microns (approx.) and then oven baked at 180 degree centigrade

Jan 11/19
PROFESSOR AND HEAD
Dept. of General Medicine
Govt. Medical College and General Hospital
Baramati, Dist- Pune

Shankar
DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE

Item No - 11

Examination Couch

SPECIFICATIONS

- A. Overall size: 1830 x 610 x 780 mm 72" x 24" x 31" ht, (L x W x H).
- B. Fixed upholstered top of 610mm, (W) / 24" (w).
- C. Optimized height (780mm) of the table for comfortable observation and reach.
- D. The design focuses on ease of operation, hygiene, doctor and patient ergonomics, and aesthetics.
Homogeneous soft forms with rounded edges evoke feeling of comfort and safety.
- E. Mattress and Upholstery is to be made with PU foam with density 40, & 75 mm thickness
- F. Detachable pillow for Headrest. 50mm thick
- G. Bottom in 3 section
- H. All part of the lower section consists of One drawer and one cupboard
- I. All the cabinets should be made from 0.8mm thick CRCA sheet and Doors to be made from 0.8 mm thick double press bent CRCA sheet with hinges. Doors should be easily detachable if needed,
- J. All doors have aesthetically designed inbuilt press bent handles placed in such a way that to give a unique look also provide wider space for gripping to users. & do not protrude outside the surface of the cupboard thus avoiding inconvenience to the patient.
- K. In built sliding step stool be made of ERW square tube , of section 25.4mm X 25.4mm and 1.25 mm thick with aluminum treads offering firm grip for climbing
- L. Under Structure and top frame is to be made of ERW square tube, of section 25.4mm X 25.4mm and 1.25 mm thick. Understructure is to be made of MS square tubes with unique styling that provides better strength and stability
- M. Finish: All components should be pretreated in separate eight tank process for better finish, good adhesion and corrosion protection. Process includes Hot Degreasing, De rusting, Activation, Phosphating & Number of Water rinses as per IS 3618 - 1966 class 'C' type and then pretreated materials is coated with epoxy powder with film thickness of 60 microns (approx.) and then oven baked at 180 degree centigrade

12/25/21
PROFESSOR AND HEAD
Dept. of General Medicine
Govt. Medical College and General Hospital
Baramati, Dist. Pune

DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE

Item no-12

SURGICAL DRUMS

SS 304 grade

Size 9" x 11" dia



DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE



PROFESSOR AND HEAD
Dept. of General Surgery
Govt. Medical College and General Hospital
Baramati, Dist- Pune

Item No - 14

AC 2 Ton

ITEM DETAILS:

- Type of Air Conditioner : High Wall Spilt AC (Hot and Cold)
- Compressor (Speed Type) : Non-inverter (Fixed Speed)
- Heating mechanism : Electric Heater
- Nominal Capacity (TR) : 2.0
- Nominal Cooling Capacity : 4500 For 2.0 TR (kcl/hr)
- Nominal Heating Capacity 90 to 100 % of Cooling Capacity
- Coil Material (Evaporator) Copper
- Coil Material (Condensor) Copper
- Fin Material (Evaporator) Copper
- Fin Material (Condensor) Copper
- Eco Friendly refrigerant : yes
- Bee Star Rating : 2
- Minimum length of Copper : NA' For Window AC/Casstte AC/Floor Standing Spilt AC
- Pipe and Suitable Connecting
- Electric Cable for Installation and commission-inclusive in the scope of supply

N.S. Adke
PROFESSOR AND HEAD
Dept. of Anesthesiology
Govt. Medical College and
General Hospital Baramati, Dist-Pune

[Signature]
DEAN
GOVERNMENT MEDICAL COLLEGE
AND GENERAL HOSPITAL
BARAMATI, DISTRICT-PUNE